

BIOCOMPATIBILITY CERTIFICATE

Testmaterial: PA 2200

Supplier: EOS GmbH
Pasinger Strasse 2, D-82152 Planegg

Studies performed: The following studies were performed in order to determine the biocompatibility of the product PA 2200 according to ISO 10993-1:

CYTOTOXICITY

SENSITISATION, polar extract

SENSITISATION, non-polar extract


INTRACUTANEOUS REACTIVITY

Results: The product did not show any adverse effects in the studies performed. Therefore, the biocompatibility of the test material was proved.

BSL BIOSERVICE Scientific Laboratories GmbH Munich

Behringstraße 6

D-82152 Planegg



Dr. Achim Albrecht

Biological Safety Testing

Date: April 10, 2001

