



BIOCOMPATIBILITY CERTIFICATE

Testmaterial: PA 2200

Supplier: EOS GmbH

Pasinger Strasse 2, D-82152 Planegg

Studies performed: The following studies were performed in order to determine the

biocompatiblity of the product PA 2200 according to ISO 10993-1:

CYTOTOXICITY

SENSITISATION, polar extract

SENSITISATION, non-polar extract

INTRACUTANEOUS REACTIVITY

Results: The product did not show any adverse effects in the studies

performed. Therefore, the biocompatiblity of the test material

was proved.

BSL BIOSERVICE Scientific Laboratories GmbH Munich

Behringstraße 6

D-82152 Planegg

Dr. Achim Albrecht

Biological Safety Testing

Date: April 10, 2001



robotmech Stössl GmbH Bundesstraße 11 A-6842 Koblach Telefon +43 5523 51031-0 Fax +43 5523 51031-80 info@robotmech.com www.robotmech.com